

EN English

Introduction

Thank you for purchasing a NC200 Mesh Nebulizer. With proper care and use, your nebulizer will provide you with many years of reliable treatments. Treatments are delivered quickly, safely and conveniently making this unit ideal for all ages. We encourage you to thoroughly read this guidebook to learn about the features of your nebulizer. Your mesh nebulizer should be used under the supervision of a licensed physician and/or a respiratory therapist; however, if you have any questions, please contact your physician or respiratory therapist to determine what are the most appropriate treatments for your respiratory condition.

NOTE: Your nebulizer is a medical device. It is not a toy. If a aerosol medication is required during therapy, Please contact with your physician and/or pharmacist to determine if your prescription medication is approved for use with the nebulizer. For type, dose, and regime of medication follow the instructions of your doctor or respiratory therapist.

This device fulfills all requirements of the EU Directive 93/42/EEC (Medical Device Directive) and the European Standard EN 13544-1:2007+A1:2009 Respiratory therapy equipment – Part 1: Nebulizers and their components.

Please read this manual carefully before use and be sure to keep this manual.

Cautions

Please use general safety precautions when operating your nebulizer. This unit should be used only for its intended purpose as described in this guidebook and with medications only under the supervision and instruction of your physician. Do not use the device in an anaesthetic or ventilator breathing circuit.

Product information

READ THIS FOLLOWING BEFORE USING:

• To avoid electrical shock, keep unit away from water.

• Do not immerse the unit in liquid.

• Do not use the unit if any part has fallen into water immediately unplug the unit.

• Do not use the unit if it has any damaged parts (including plug), if it has been submerged in water or dropped. Remove the unit from the water and repair or replace.

• Dry clean the unit before use.

• Do not touch the mesh with any other material.

• Do not use the unit if it is completely dry.

• Ensure that no water penetrates into the main unit. Any liquid that penetrates into the main unit can damage the electrical parts or other nebulizer components and impair the function of the device.

Cleaning the unit

Wipe the unit daily using a damp lint free cloth and use a dry lint free cloth to wipe the main unit.

NOTE: Never clean with benzene or thinner. Any other form of cleaning or cleaning agents may damage the finish of the unit.

Troubleshooting

PROBLEM

Check the AC Adapter (optional). Is the plug of the Adapter into a working outlet while working or not?

The power indicator does not illuminate. Check the batteries are inserted into the main unit properly.

The power indicator is flashing in green light. Check the power of the batteries is empty or not.

The power indicator is flashing in red light. Check the connectors between the main unit and medication cup to see if they are dirty or not.

The power indicator is shut down immediately after the device is turned on. Check the medication cup is already inserted into the main unit or not.

Malfuction. Return the device to your local dealer or service center.

The power indicator is not visible aerosol. Fill the medication cup with pre-specified amount of medication and then turn on the device.

The power indicator is illuminated but no visible aerosol. Remove the air bubble by shaking the medication cup refer to G-1.

Check the medication cup to see if it is dirty or not. Check the medication cup refer to "Cleaning procedures".

Excessive medication is collected on the mesh (like droplet) outside the medication cup. Replace the medication cup with new one. (refer to G-1)

The medication is not suitable for spraying (e.g. too thick). Remove the air bubble by shaking the medication cup.

The power indicator is illuminated but no visible aerosol. Make sure that the nebulizer is not held in an upright position. Make sure that the medication cup is filled with 2-6 ml amount of medicine.

The medication is not suitable for spraying (e.g. too thick). Make sure that the medication cup is filled with 2-6 ml amount of medicine.

The power indicator is illuminated but no visible aerosol. Too little medicine in the nebulizer. Remove the air bubble by shaking the medication cup.

The medication is not suitable for spraying (e.g. too thick). Remove the air bubble by shaking the medication cup.

The power indicator is illuminated but no visible aerosol. Inspect the medication cup to see if it is dirty or not.

The power indicator is illuminated but no visible aerosol. Clean the Medication cup refer to "Cleaning procedures".

The power indicator is illuminated but no visible aerosol. Replace the medication cup with new one.

If above solutions do not work, call your local supplier for further help.

Product specifications

Battery: 3V AA Alkaline*2)

Power Consumption: <1.2W

Sound Level: <40dB(A) 1 meter away from NC200

Operation frequency: Approx. 120 kHz

Operating Temperature Range: 10°C to 40°C (50°F to 104°F)

Operating Humidity Range: 20% to 100% RH

Storage Temperature Range: -10°C to 60°C (14°F to 140°F)

Storage Humidity Range: ≤ 90% RH

Dimension (L x W x H): 42 x 70 x 117 mm (16.5 x 27.6 x 46.1 inches) (without batteries and accessories)

Medication Capacity: Max. 8 ml/cup

Average Nebulization Rate: ≤ 5 ml/min (0.96 Saline Solution)

Standard Accessories: Medication Cup, Mouthpiece, Child mask, USB cable

Subject to technical modification without prior notice.

Performance may vary with drugs such as suspensions or high viscosity. See drug information for further details.

Cleaning cautions

Check nebulizer, mouthpiece and any other optional component before each use. Dirty or worn parts should not be replaced.

Do not wash the unit in water. It may damage the unit.

Do not use this unit while operating a ventilator.

Pentamidine is not an approved medication for this device.

Stoppage: Turn off the unit in direct sunlight, high temperature or humidity.

Keep the unit out of reach of children.

Always keep the unit unplugged while not in use.

Cleaning caution

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

Disconnect the unit from the electrical outlet before cleaning.

Clean all necessary parts after each use as instructed in this guidebook.

Always dispose of any remaining medication in the medication cup after each use. Use the disposal bin provided with the unit to store the device.

Do not use gravity atomizers. CO2 gas dislocation or low temperature plasma sterilizer to disinfect the unit and parts.

Only clean the nebulizer after every application for maximum hygiene to avoid cross-contamination.

</